



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

MEYER GAGE COMPANY
230 Burnham Street
South Windsor, CT 06074
Melissa Lee Phone: 860 528 6526
quality@mevergage.com

CALIBRATION

Valid until: August 31, 2014

Certificate Number: 3340.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2,3} (±)	Comments
Cylindrical Plug Gages – Class X & XX	Up to 4.000 in	(6.0 + 1.9L) µin	P&W Labmaster & gage blocks
Class Z & ZZ	(0.005 to 1.000) in	24 µin	Laser micrometer
Cylindrical Ring Gages – Class X & XX	(1.000 to 4.000) in	(8.0 + 3.0L) µin	P&W Labmaster & master ring gages

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ In the statement of CMC, L is the numerical value of the nominal length of the device measured in inches.

A handwritten signature in black ink, appearing to read "Peter Abney". The signature is written in a cursive style with a large, sweeping initial "P".



The American Association for Laboratory Accreditation

World Class Accreditation

Accredited Laboratory

A2LA has accredited

MEYER GAGE COMPANY INC.

South Windsor, CT

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).

Presented this 18th day of July 2012.





Peter Meyer

President & CEO
For the Accreditation Council
Certificate Number 3340.01
Valid to August 31, 2014

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.