



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

MEYER GAGE COMPANY INC.  
 230 Burnham Street  
 South Windsor, CT 06074  
 Melissa Lee Phone: 860 528 6526  
[quality@mebergage.com](mailto:quality@mebergage.com)

CALIBRATION

Valid To: August 31, 2022

Certificate Number: 3340.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments	
Cylindrical Plug Gages and Discs – External Diameter	Class ZZ, Z, Y, X & XX	8.3 µin	P&W Labmaster™ & gage blocks	
	Class ZZ & Z	11.0 µin + 0.80 µin/in		
		(0.005 to 1.015) in	28 µin	Laser micrometer
Cylindrical Ring Gages – Internal Diameter	Class Z, Y, X & XX	9.7 µin + 2.4 µin/in	P&W Labmaster™ & master ring gages	
		(0.125 to 1) in		9.9 µin + 1.1 µin/in
		(1 to 4) in		11 µin

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Thread Wires	(1 to 120) TPI (0.2 to 10) Pitch	7.2 μin 0.20 μm	P&W Labmaster™ & master gage blocks
Gear Wires	(2 to 200) P	8.3 μin	P&W Labmaster™ & master gage blocks

---

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.



# Accredited Laboratory

A2LA has accredited

**MEYER GAGE COMPANY INC.**

*South Windsor, CT*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 30<sup>th</sup> day of November 2020.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 3340.01  
Valid to August 31, 2022

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*